

Hospital Grade Power Cords

QUALTEK ELECTRONICS

Hospital Grade - Unshielded with North American Color Code



I



J



K



L

Part Number	Figure	Length	Cordage	# Cond. AWG	Amp Rating	Plug No.	Conn No.
232002-06	I	8' 0"	SJT+	3 x 18	10	019*	--
232003-06	I	10' 0"	SJT+	3 x 18	10	019*	--
232010-06	I	8' 0"	SJT+	3 x 18	10	019**	--
232011-06	I	10' 0"	SJT+	3 x 18	10	019**	--
233002-06	J	8' 0"	SJT+	3 x 18	10	019*	005N
233003-06	J	10' 0"	SJT+	3 x 18	10	019*	005N
233008-06	J	8' 0"	SJT+	3 x 18	10	019**	005N
233009-06	J	10' 0"	SJT+	3 x 18	10	019**	005N
242003-06	K	10' 0"	SJT+	3 x 18	10	019L*	--
243003-06	L	10' 0"	SJT+	3 x 18	10	019L*	005N

* Hospital Grade Clear Plug

** Hospital Grade Gray Plug

+ North American Color Code (Black/White/Green)

All Cords have 60°C and 125VAC unless marked as above

Qualtek Electronics *Meeting Global Demands with Design Solutions*

7675 Jenther Drive ☐ Mentor, Ohio 44060 ☐ (888) 258-3468 ☐ (440) 951-3300 ☐ FAX: (440) 951-7252

E-Mail mailbox@qualtekusa.com ☐ Web: www.qualtekusa.com

Copyright © 2004 Qualtek Electronics Corporation. All rights reserved.

Updated: March, 2004